

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 27 2011

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130302

Name CHRIS HILL

ID# 3839 Date 7/18/11

A Agency AST

Phone # 486-4121

Instrument Location KODIAK POST, INTERVIEW ROOM

B Alco S/N X124658 Target Value .084 High Pressure 1140

Alco Test Value Average .086 .087
1st Alco 2nd Alco

Signature [Signature]

(OVER)

(Do Not write in the area below)

MBB
7/27/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug, 2010. NJB 2011

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas NIKKI ROTH
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

JULY 18, 2011

OPERATOR'S NAME:

HILL C W

OPERATOR'S NUMBER: 5620

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME: M

X

O.L. #: 0

DEPT/AGENCY: PDH

CASE/REPORT: 0

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: X124653

--- BREATH ANALYSTS ---

.084 ADJUSTED FOR 30.04 in
ALCO TARGET .084 08:53
BLANK TEST .000 08:54
INTERNAL STANDARD VERIFIED 08:54
ALCO TV 30.04 in .086 08:55
BLANK TEST .000 08:55
SUBJECT SAMPLE .084 08:56
BLANK TEST .000 08:57
ALCO TV 30.04 in .087 08:57
BLANK TEST .000 08:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

JULY 18, 2011

TIME 08:50

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/09/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.04 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~